# PRODUCT SPECIFICATION



#### **PRODUCT INFORMATION**

Trade Name MEMORIAL GLOVES

Intended Use Worn on examiner's hand or finger to prevent

contamination between patient and examiner.

Materials 100% synthetic nitrile

Color Blue.

Type Nitrile Examination Glove, Powder Free, Polymer

-coated, Non-sterile, Ambidextrous.

Surface Finish Finger textured, beaded cuff.

Powder No powder lubricant added.

Storage Condition The gloves maintain their properties when stored

in a dry condition. Avoid direct sunlight.

Shelf Life The gloves have shelf life of 3 years from the date

of manufacture with the above storage conditions.

100 !

UOM 100 gloves.

Sizes Small, Medium, Large, X Large



## **PRODUCT INFORMATION**

DIMENSIONS	SIZE	LENGTH (MM) MIN	WIDTH (MM)
	SMALL	>230	70 ± 10
	MEDIUM	>230	80 ± 10
	LARGE	>230	95 ± 10
EX	(TRA- LARGE	>230	110 ± 10

## **PACKAGING CONFIGURATION & LABELLING:**

Packaging Style: 100 gloves X 10 boxes X 1 carton

Size Marking: The size of the gloves shall be marked in the

check box on every carton with black ink.

Packing: 100 gloves per inner box (by count) 10 boxes

per outer carton

Labelling: The inner boxes and outer cartons have the

following information: product description, brand, size and net quantity of gloves, country of origin. name and address of distributor, and

Aged

>14

>400

single use only symbol.

**Un-aged** 

>14

>500

Barcodes: GS1 format.

Thickness	Palm	Finger	Force at Break
Millimetres	>0.08mm	>0.08mm	MPa
(MILS)	(>4.0)	(>4.0)	Elongation (%)

# **QUALITY SPECIFICATION\***

#### FOR MORE INFORMATION:

**300 Commerce Drive** 

Lackawanna, NY 14218

info@glovesupply.us

(212) 771-8822

**CONTACT US** 

www.glovesupply.us

INSPECTION CRITERIA SAMPLING PLAN/ AQL

Dimension S2 - AQL 4.0

Physical Properties S2 - AQL 4.0

Water Tight test G1 - AQL 1.5

Visual Inspection - major G1 - AQL 2.5

Visual Inspection - minor G1 - AQL 4.0

Residual Powder n = 5 (average)

Mix size/ mix glove/ mix hand Not allowed