# PRODUCT SPECIFICATION



#### PRODUCT INFORMATION

Trade Name MEMORIAL GLOVES

Intended Use Worn on examiner's hand or finger to prevent

contamination between patient and examiner.

Materials 100% synthetic nitrile

Color Blue.

Type Nitrile Examination Glove, Powder Free, Polymer

-coated, Non-sterile, Ambidextrous.

Surface Finish Finger textured, beaded cuff.

Powder No powder lubricant added.

Storage Condition The gloves maintain their properties when stored

in a dry condition. Avoid direct sunlight.

Shelf Life The gloves have shelf life of 3 years from the date

of manufacture with the above storage conditions.

UOM 100 gloves.

Sizes Small, Medium, Large, X Large



#### PRODUCT INFORMATION

Dimensions	SIZE	LENGTH (MM) MIN	WIDTH (MM)
Millimetres	SMALL	>230	70 ± 10
	MEDIUM	>230	80 ± 10
	LARGE	>230	95 ± 10
	EXTRA- LARGE	>230	110 + 10

#### PACKAGING CONFIGURATION & LABELLING:

Packaging Style: 100 gloves X 10 boxes X 1 carton

Size Marking: The size of the gloves are marked in the

(check box) on every carton with black ink.

Packing: 100 gloves per inner box (by count) 10 boxes

per outer carton

Labelling: The inner boxes and outer cartons have the

following information: product description, brand, size and net quantity of gloves, country of origin, name and address of distributor, and

single use only symbol.

Barcodes: GS1 format.

Thickness	Palm	Finger	Force at Break	Un-aged	Aged
Millimetres (MILS)	<b>&gt;0.10mm</b> (>4.0)	<b>&gt;0.10mm</b> (>4.0)	MPa Elongation (%)	<b>&gt;14</b> >500	<b>&gt;14</b> >400

## CONTACT US

### FOR MORE INFORMATION:

1 Linden Place, Suite 303 Great Neck, NY 11021

info@glovesupply.us

(212) 771-8822

www.glovesupply.us

#### QUALITY SPECIFICATION<sup>®</sup>

INSPECTION CRITERIA	SAMPLING PLAN/ AQL		
Dimension	S2 - AQL 4.0		
Physical Properties	S2 - AQL 4.0		
202000000000000000000000000000000000000			

Water Tight test G1 - AQL 1.5

Visual Inspection - major G1 - AQL 2.5

Visual Inspection - minor G1 - AQL 4.0

Residual Powder n = 5 (average)

Mix size/ mix glove/ mix hand Not allowed